




Hydrocarbons (Lower Class) C₂ to C₇

Part No.: 103

Hydrocarbons (Lower Class) C₂ to C₇ No.103



When used, these tubes are to be connected. See page 2-3.

Performance

This detector tube is calibrated with Propane.

Measuring range	0.05 to 0.1 %	0.1 to 1.2 %	1.2 to 2.4 %
Number of pump strokes	2 (200 mL)	1 (100 mL)	1/2 (50 mL)
Correction factor	1/2	1	2
Sampling time	4 min	2 min	1 min
Detecting limit :	0.01 % (2 pump strokes)		
Colour change :	Yellowish brown → Greenish brown		
Operating conditions :	Temperature 0 to 40 °C (32 to 104 °F) correction not used		
	Relative humidity 0 to 90 % correction not used		
Relative standard deviation :	10 % (for 0.1 to 1.2 %)		
Tube quantity and number of tests per box :	10 tubes for 9 tests		
Shelf life :	30 months		

Reaction principle



Possible coexisting substances and their interferences

Substance	Concentration	Interference	Changes colour by itself to
Organic solvents		+	Greenish brown
Carbon monoxide		No	No
Ethane, Methane		No	No

Substances measurable with this detector tube

Substance	Correction factor	n	Measuring range (%)
Acetylene	3	1/2	1.8 to 3.6
	1.5	1	0.15 to 1.8
	0.75	2	0.075 to 0.15
Isobutane, Butane, Heptane	1.4	1/2	0.84 to 1.68
	0.7	1	0.07 to 0.84
	0.35	2	0.035 to 0.07
Isopentane	1.8	1/2	1.08 to 2.16
	0.9	1	0.09 to 1.08
	0.45	2	0.045 to 0.09
Ethylene	14	1/2	8.4 to 16.8
	7	1	0.7 to 8.4
	3.5	2	0.35 to 0.7
Propane	2	1/2	1.2 to 2.4
	1	1	0.1 to 1.2
	0.5	2	0.05 to 0.1
Hexane	1	1/2	0.6 to 1.2
	0.5	1	0.05 to 0.6
	0.25	2	0.025 to 0.05
Pentane	1.5	1/2	0.9 to 1.8
	0.75	1	0.075 to 0.9
	0.375	2	0.0375 to 0.075

Calibration gas generation

High pressure gas cylinder method

Special note

When measuring, the SO₃ filter tube (white) should be connected to the back of the detector tube. SO₃ filter tube can be reused up to nine times.
Nine detector tubes and one SO₃ filter tube are contained in one box.